



Masks with magnetic clips¹ crosswalk guide

We have compiled a list of alternative masks from our existing portfolio to assist our customers in advising patients who are contraindicated due to magnetic components. Please refer to the Urgent Field Safety Notice we previously sent you regarding the updated contraindications and warnings for ResMed masks containing magnetic clips. For compatibility information, please consult the product user guide or the mask/device compatibility list found at ResMed.com/downloads/masks.

It's important to note that sizing information is available as a reference, but performing a fitting test is critical to ensure proper mask fit. Detailed instructions and precautions can be found in the user guide.

Full face masks

Masks with Magnetic Clips ¹	
	
AirFit™ F20 / AirFit™ F20 for Her	
Mask System	
63400	AirFit F20 - SML
63400	AirFit F20 - SML
63401	AirFit F20 - MED
63402	AirFit F20 - LGE
63403	AirFit F20 for Her - SML
63403	AirFit F20 for Her - SML
63404	AirFit F20 for Her - MED

Masks without Magnetic Clips



AirFit™ F20 Non Magnetic
(Available in the U.S. Only)



AirFit™ F10 /
AirFit™ F10 for Her



Mirage Quattro™

Mask System

64028	AirFit F20 NM - SML	63101	AirFit F10 - SML	61200	Mirage Quattro - X-SML
64028	AirFit F20 NM - SML	63101	AirFit F10 - SML	61201	Mirage Quattro - SML
64029	AirFit F20 NM - MED	63102	AirFit F10 - MED	61202	Mirage Quattro - MED
64030	AirFit F20 NM - LGE	63103	AirFit F10 - LGE	61203	Mirage Quattro - LGE
64028	AirFit F20 NM - SML	63139	AirFit F10 for Her - X-SML	61200	Mirage Quattro - X-SML
64028	AirFit F20 NM - SML	63140	AirFit F10 for Her - SML	61201	Mirage Quattro - SML
64029	AirFit F20 NM - MED	63141	AirFit F10 for Her - MED	61202	Mirage Quattro - MED

*Nasal pillows masks without magnetic clips include AirFit™ P10 and AirFit™ P30i.
Please visit ResMed.com for more information regarding these masks.*



Masks with magnetic clips¹ crosswalk guide

We have compiled a list of alternative masks from our existing portfolio to assist our customers in advising patients who are contraindicated due to magnetic components. Please refer to the Urgent Field Safety Notice we previously sent you regarding the updated contraindications and warnings for ResMed masks containing magnetic clips. For compatibility information, please consult the product user guide or the mask/device compatibility list found at ResMed.com/downloads/masks.

It's important to note that sizing information is available as a reference, but performing a fitting test is critical to ensure proper mask fit. Detailed instructions and precautions can be found in the user guide.

Full face masks

Masks with Magnetic Clips ¹	
	
AirTouch™ F20 / AirTouch™ F20 for Her	
Mask System	
63000	AirTouch F20 - SML
63000	AirTouch F20 - SML
63001	AirTouch F20 - MED
63002	AirTouch F20 - LGE
63003	AirTouch F20 for Her - SML
63003	AirTouch F20 for Her - SML
63004	AirTouch F20 for Her - MED

Masks without Magnetic Clips



AirFit™ F20 Non Magnetic
(Available in the U.S. Only)



AirFit™ F10 /
AirFit™ F10 for Her



Mirage Quattro™

Mask System

64028	AirFit F20 NM - SML	63101	AirFit F10 - SML	61200	Mirage Quattro - X-SML
64028	AirFit F20 NM - SML	63101	AirFit F10 - SML	61201	Mirage Quattro - SML
64029	AirFit F20 NM - MED	63102	AirFit F10 - MED	61202	Mirage Quattro - MED
64030	AirFit F20 NM - LGE	63103	AirFit F10 - LGE	61203	Mirage Quattro - LGE
64028	AirFit F20 NM - SML	63139	AirFit F10 for Her - X-SML	61200	Mirage Quattro - X-SML
64028	AirFit F20 NM - SML	63140	AirFit F10 for Her - SML	61201	Mirage Quattro - SML
64029	AirFit F20 NM - MED	63141	AirFit F10 for Her - MED	61202	Mirage Quattro - MED

*Nasal pillows masks without magnetic clips include AirFit™ P10 and AirFit™ P30i.
Please visit ResMed.com for more information regarding these masks.*



Masks with magnetic clips¹ crosswalk guide

We have compiled a list of alternative masks from our existing portfolio to assist our customers in advising patients who are contraindicated due to magnetic components. Please refer to the Urgent Field Safety Notice we previously sent you regarding the updated contraindications and warnings for ResMed masks containing magnetic clips. For compatibility information, please consult the product user guide or the mask/device compatibility list found at ResMed.com/downloads/masks.

It's important to note that sizing information is available as a reference, but performing a fitting test is critical to ensure proper mask fit. Detailed instructions and precautions can be found in the user guide.

Full face masks

Masks with Magnetic Clips ¹	
	
AirFit™ F30	
Mask System	
64100	AirFit F30 - SML
64100	AirFit F30 - SML
64101	AirFit F30 - MED
64101	AirFit F30 - MED

Masks without Magnetic Clips



AirFit™ F20 Non Magnetic

(Available in the U.S. Only)



AirFit™ F10 /

AirFit™ F10 for Her



Mirage Quattro™

Mask System

64028	AirFit F20 NM - SML	63101	AirFit F10 - SML	61200	Mirage Quattro - X-SML
64028	AirFit F20 NM - SML	63101	AirFit F10 - SML	61201	Mirage Quattro - SML
64029	AirFit F20 NM - MED	63102	AirFit F10 - MED	61202	Mirage Quattro - MED
64030	AirFit F20 NM - LGE	63103	AirFit F10 - LGE	61203	Mirage Quattro - LGE

Nasal pillows masks without magnetic clips include AirFit™ P10 and AirFit™ P30i.
Please visit ResMed.com for more information regarding these masks.



Masks with magnetic clips¹ crosswalk guide

We have compiled a list of alternative masks from our existing portfolio to assist our customers in advising patients who are contraindicated due to magnetic components. Please refer to the Urgent Field Safety Notice we previously sent you regarding the updated contraindications and warnings for ResMed masks containing magnetic clips. For compatibility information, please consult the product user guide or the mask/device compatibility list found at ResMed.com/downloads/masks.

It's important to note that sizing information is available as a reference, but performing a fitting test is critical to ensure proper mask fit. Detailed instructions and precautions can be found in the user guide.

Full face masks

Masks with Magnetic Clips ¹	
	
AirFit™ F30i	
Mask System	
63330	AirFit F30i - S/SML
63330	AirFit F30i - S/SML
63331	AirFit F30i - S/STD
63332	AirFit F30i - M/STD
63334	AirFit F30i - M/LGE
63333	AirFit F30i - W/STD

Masks without Magnetic Clips



AirFit™ F20 Non Magnetic
(Available in the U.S. Only)



AirFit™ F10 /
AirFit™ F10 for Her



Mirage Quattro™

Mask System

64028	AirFit F20 NM - SML	63101	AirFit F10 - SML	61200	Mirage Quattro - X-SML
64028	AirFit F20 NM - SML	63101	AirFit F10 - SML	61201	Mirage Quattro - SML
64028	AirFit F20 NM - SML	63101	AirFit F10 - SML	61201	Mirage Quattro - SML
64029	AirFit F20 NM - MED	63102	AirFit F10 - MED	61202	Mirage Quattro - MED
64029	AirFit F20 NM - MED	63102	AirFit F10 - MED	61202	Mirage Quattro - MED
64030	AirFit F20 NM - LGE	63103	AirFit F10 - LGE	61203	Mirage Quattro - LGE

Nasal pillows masks without magnetic clips include AirFit™ P10 and AirFit™ P30i.
Please visit ResMed.com for more information regarding these masks.



Masks with magnetic clips¹ crosswalk guide

We have compiled a list of alternative masks from our existing portfolio to assist our customers in advising patients who are contraindicated due to magnetic components. Please refer to the Urgent Field Safety Notice we previously sent you regarding the updated contraindications and warnings for ResMed masks containing magnetic clips. For compatibility information, please consult the product user guide or the mask/device compatibility list found at [ResMed.com/downloads/masks](https://resmed.com/downloads/masks).

It's important to note that sizing information is available as a reference, but performing a fitting test is critical to ensure proper mask fit. Detailed instructions and precautions can be found in the user guide.

Nasal masks

Masks with Magnetic Clips ¹	
	
AirFit™ N20 / AirFit™ N20 for Her	
Mask System	
63503	AirFit N20 - SML
63501	AirFit N20 - MED
63502	AirFit N20 - LGE
63502	AirFit N20 - LGE
63500	AirFit N20 for Her

Masks without Magnetic Clips			
			
AirFit™ N30		AirFit™ N30i²	
Mask System			
64222	AirFit N30 - SML	63801	AirFit N30i STPK - SML
64223	AirFit N30 - MED	63800	AirFit N30i STPK - STD
64223	AirFit N30 - MED	63800	AirFit N30i STPK - STD
64224	AirFit N30 - SML-WIDE	63800	AirFit N30i STPK - STD
64222	AirFit N30 - SML	63801	AirFit N30i STPK - SML

Nasal pillows masks without magnetic clips include AirFit™ P10 and AirFit™ P30i. Please visit [ResMed.com](https://resmed.com) for more information regarding these masks.



Masks with magnetic clips¹ crosswalk guide

We have compiled a list of alternative masks from our existing portfolio to assist our customers in advising patients who are contraindicated due to magnetic components. Please refer to the Urgent Field Safety Notice we previously sent you regarding the updated contraindications and warnings for ResMed masks containing magnetic clips. For compatibility information, please consult the product user guide or the mask/device compatibility list found at [ResMed.com/downloads/masks](https://resmed.com/downloads/masks).

It's important to note that sizing information is available as a reference, but performing a fitting test is critical to ensure proper mask fit. Detailed instructions and precautions can be found in the user guide.

Nasal masks

Masks with Magnetic Clips ¹	
	
AirTouch™ N20 / AirTouch™ N20 for Her	
Mask System	
63903	AirTouch N20 - SML
63901	AirTouch N20 - MED
63902	AirTouch N20 - LGE
63902	AirTouch N20 - LGE
63900	AirTouch N20 for Her

Masks without Magnetic Clips			
			
AirFit™ N30		AirFit™ N30i ²	
Mask System			
64222	AirFit N30 - SML	63801	AirFit N30i STPK - SML
64223	AirFit N30 - MED	63800	AirFit N30i STPK - STD
64223	AirFit N30 - MED	63800	AirFit N30i STPK - STD
64224	AirFit N30 - SML-WIDE	63800	AirFit N30i STPK - STD
64222	AirFit N30 - SML	63801	AirFit N30i STPK - SML

Nasal pillows masks without magnetic clips include AirFit™ P10 and AirFit™ P30i. Please visit [ResMed.com](https://resmed.com) for more information regarding these masks.



Masks with magnetic clips¹ crosswalk guide

We have compiled a list of alternative masks from our existing portfolio to assist our customers in advising patients who are contraindicated due to magnetic components. Please refer to the Urgent Field Safety Notice we previously sent you regarding the updated contraindications and warnings for ResMed masks containing magnetic clips. For compatibility information, please consult the product user guide or the mask/device compatibility list found at [ResMed.com/downloads/masks](https://resmed.com/downloads/masks).

It's important to note that sizing information is available as a reference, but performing a fitting test is critical to ensure proper mask fit. Detailed instructions and precautions can be found in the user guide.

Nasal masks

Masks with Magnetic Clips ¹	
	
AirFit™ N10 / AirFit™ N10 for Her	
Mask System	
63229	AirFit N10 - SML
63200	AirFit N10 - STD
63202	AirFit N10 - WIDE
63202	AirFit N10 - WIDE
63201	AirFit N10 for Her

Masks without Magnetic Clips			
			
AirFit™ N30		AirFit™ N30i ²	
Mask System			
64222	AirFit N30 - SML	63801	AirFit N30i STPK - SML
64223	AirFit N30 - MED	63800	AirFit N30i STPK - STD
64223	AirFit N30 - MED	63800	AirFit N30i STPK - STD
64224	AirFit N30 - SML-WIDE	63800	AirFit N30i STPK - STD
64222	AirFit N30 - SML	63801	AirFit N30i STPK - SML

¹Nasal pillows masks without magnetic clips include AirFit™ P10 and AirFit™ P30i. Please visit [ResMed.com](https://resmed.com) for more information regarding these masks.



This content is intended for health professionals only. Please refer to the user guide for relevant information related to any warnings and precautions to be considered before and during use of the product.

1. The mask contains magnets that may interfere with certain implants or medical devices. They are not to be used by patients where they, or anyone in close physical contact while using the mask, have active medical implants that interact with magnets and/or metallic implants/objects containing ferromagnetic material. Keep the mask magnets at a safe distance of at least 6 inches (150 mm) away from implants or medical devices that may be adversely affected by magnetic interference. Please refer to the User Guide for complete information, including magnet contraindications and warnings.
2. AirFit™ N30i Starter Packs (both Small and Standard) are supplied with Small, Medium, and Small-Wide cushion sizes.